




















2

Assay Data Sheet - Expected Ranges BioProfile® FLEX Controls - LEVEL 2

2

Lot Number:  712034				Exp. Date: 2009-06	
Expected Range				+/- 3 SD Expected Range	
Parameter	Units	Lower Limit	Upper Limit	Lower Limit	Upper Limit
pH	-	 6.875	 6.975	 6.850	 7.000
PO ₂	mm Hg	 99.2	 119.2	 94.2	 124.2
PCO ₂	mm Hg	 39.9	 48.7	 37.8	 50.8
Gluc	g/L	 4.25	 5.75	 3.87	 6.13
Lac	g/L	 2.55	 3.45	 2.32	 3.68
NH ₄ ⁺	mmol/L	 3.85	 5.65	 3.35	 6.15
Na ⁺	mmol/L	 122.5	 140.9		
K ⁺	mmol/L	 3.58	 4.42		
Ca ⁺⁺	mmol/L	 0.67	 0.93		
Osmolality	mOsm/kg H ₂ O	 391	 403		

nova
biomedical

200 Prospect Street • Waltham, MA U.S.A. 02454-9141 • (781) 864-0800

LFN 08240A 2007-08