



CERTIFICATE OF ANALYSIS

STP pHox Plus L Cal Cartridge
Lot Number 16095034

Ref Number 34088
Exp Date 2017-09-16
Date of Manufacture 2016-03-16

Cal A	L/N	16076051	Cal B	L/N	16076051	Cal C	L/N	16076051
	Spec	Pass/Fail		Spec	Pass/Fail		Spec	Pass/Fail
Na mmol/L	142-144	Pass	HCO3 mmol/L	45.90-47.78	Pass	pH @ 37C	7.332-7.338	Pass
HCO3 mmol/L	24.42-25.42	Pass				Ca mmol/L	2.01-2.07	Pass
K mmol/L	4.03-4.09	Pass				Na mmol/L	72.5-73.5	Pass
Cl mmol/L	97.2-98.8	Pass				K mmol/L	10.04-10.20	Pass
Ca mmol/L	1.01-1.03	Pass				Cl mmol/L	52.6-53.4	Pass
Glu mg/dL	39.6-40.4	Pass				Glu mg/dL	198-202	Pass
Lac mmol/L	0.8-1.2	Pass				Lac mmol/L	5.5-6.5	Pass
Cal D	L/N	16091009	Reference	L/N	16076051			
	Spec	Pass/Fail		Spec	Pass/Fail			
pH @ 37C	6.788-6.794	Pass	KCl	1.8-2.2	Pass			
Glu mg/dL	990-1010	Pass						
Lac mmol/L	10.5-13.5	Pass						

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U. S. A. in accordance with 21 CFR-Part 820, Quality System Regulation (QSR), Current Good Manufacturing Practice (cGMP) and EN ISO 13485:2012, Medical Devices, Quality Management Systems for in Vitro Diagnostic devices, and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Approval

QC Supervisor
Title

4/15/2016
Date