



### CERTIFICATE OF ANALYSIS

STP CCX CO-Oximeter Reagent Cartridge  
Lot Number 16342019

Ref Number 35150  
Exp Date 2017-12-07  
Date of Manufacture 2016-12-07

Dil L/N	16342019		Flush L/N	16342020	
	Spec	Result		Spec	Result
Surfactant	N/A	Present	Surfactant	N/A	Present
Buffers	N/A	Present	Preservatives	N/A	Present
Preservatives	N/A	Present			

THb Calibrator 16342020  
g/dL 16.1  
g/L 161  
mmol/L 10.0

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U. S. A. in accordance with 21 CFR-Part 820, Quality System Regulation (QSR), Current Good Manufacturing Practice (cGMP) and EN ISO 13485:2012, Medical Devices, Quality Management Systems for in Vitro Diagnostic devices, and conforms to the indicated test specifications. Total Hemoglobin (THb) is traceable by using Cyanmethemoglobin method.

Approval

QC Supervisor

Title

2/14/2017

Date