



CERTIFICATE OF ANALYSIS

STP CCX Cal Cartridge 4
Lot Number 16246058

Ref Number 37164
Exp Date 2017-08-19
Date of Manufacture 2016-08-16

Cal A	L/N	16235008	Cal B	L/N	16244029	Cal C	L/N	16235012
	Spec	Pass/Fail		Spec	Pass/Fail		Spec	Pass/Fail
pH @ 37C	7.351-7.357	Pass	pH @ 37C	6.822-6.828	Pass	Na mmol/L	139.0-141.0	Pass
Lac mmol/L	13.5-16.5	Pass	Na mmol/L	69.5-70.5	Pass	K mmol/L	3.97-4.03	Pass
			K mmol/L	9.9-10.1	Pass	Cl mmol/L	118.0-120.0	Pass
			Cl mmol/L	39.7-40.3	Pass	BUN mg/dl	9.9-10.1	Pass
			Ca mmol/L	1.97-2.03	Pass	Glu mg/dl	78.8-81.2	Pass
			Mg mmol/L	1.48-1.52	Pass	Lac mmol/L	1.8-2.2	Pass
						Ca mmol/L	0.98-1.02	Pass
						Mg mmol/L	0.49-0.51	Pass
						HCO3 mmol/L	18.43-19.19	Pass
Cal D	L/N	16235013	Flush	L/N	16238034	Reference	L/N	16232020
	Spec	Pass/Fail		Spec	Result		Spec	Pass/Fail
BUN mg/dl	49.4-50.6	Pass	Aqueous Salt	N/A	Present	KCl mol/L	1.8-2.2	Pass
Glu mg/dl	197-203	Pass	Solution					
Lac mmol/L	5.5-6.5	Pass						
HCO3 mmol/L	37.05-38.57	Pass						

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U. S. A. in accordance with 21 CFR-Part 820, Quality System Regulation (QSR), Current Good Manufacturing Practice (cGMP) and EN ISO 13485:2012, Medical Devices, Quality Management Systems for in Vitro Diagnostic devices, and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Approval

QC Supervisor

Title

9/16/2016

Date