




CERTIFICATE OF ANALYSIS

Bioprofile 100 Plus Reagent Pack
Lot Number 619013

Part Number 38291
Exp Date 10/31/2017

Std A	L/N	16190013		Std B	L/N	16190013		Std C	L/N	16190013	
		Spec	Result			Spec	Result			Spec	Result
pH @ 37C		7.381-7.387	Pass	pH @ 37C		6.837-6.843	Pass	Na mmol/L		139-141	Pass
Lac g/L		4.5-5.5	Pass					K mmol/L		3.97-4.03	Pass
Glu mmol/L		4.5-5.5	Pass					Lac g/L		0.9-1.1	Pass
Gluc g/L		14.3-15.7	Pass					Gluc g/L		0.9-1.1	Pass
								NH4 mmol/L		0.99-1.01	Pass
								Gln mmol/L		0.90-1.10	Pass
								Glu mmol/L		0.4-0.6	Pass
Std D	L/N	16190013		Flush	L/N	16190013		Reference	L/N	16190013	
		Spec	Result			Spec	Result			Spec	Result
Na mmol/L		74.5-75.5	Pass	Aqueous Salt		N/A	Present	KCl mol/L		1.8-2.2	Pass
K mmol/L		9.92-10.08	Pass	Solution							
Lac g/L		2.7-3.3	Pass								
Gluc g/L		4.5-5.5	Pass								
NH4 mmol/L		4.95-5.05	Pass								
Gln mmol/L		4.50-5.50	Pass								
Glu mmol/L		1.75-2.25	Pass								

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U. S. A. in accordance with ISO 9001:2008, Quality Management System Requirements and current Good Manufacturing Practices (cGMP) , and conforms to the indicated test specifications. The following analytes are traceable to NIST SRM Materials: pH, Lac, Gluc, Na, K, NH4, KCl.


Approval

QC Supervisor
Title

7/22/2016
Date