



CERTIFICATE OF ANALYSIS

BioProfile FLEX Nutrient/Metabolite Cal Cartridge
Lot Number 16183008

Part Number 39544
Exp Date 2017-09

CAL N1 L/N 16182024

CAL N2 L/N 16183007

Nutrient Flush L/N 16183008

	Spec	Result
Gln mmol/L	0.09-0.11	Pass
Glu mmol/L	0.045-0.055	Pass
Gluc g/L	0.09-0.11	Pass
Lac g/L	0.09-0.11	Pass

	Spec	Result
Gln mmol/L	0.45-0.55	Pass
Glu mmol/L	0.175-0.225	Pass
Gluc g/L	0.45-0.55	Pass
Lac g/L	0.27-0.33	Pass

	Spec	Result
pH	6.900-7.100	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U. S. A. in accordance with ISO 9001:2008, Quality Management System Requirements and current Good Manufacturing Practices (cGMP) , and conforms to the indicated test specifications. The following analytes are traceable to NIST SRM Materials: Gluc, Lac.

Approval

QC Supervisor

Title

7/13/2016

Date