



### CERTIFICATE OF ANALYSIS

Bio pHox Cal Pack

Part Number 41874

Lot Number 16098049

Exp Date 9/30/2017

Std A	L/N	16091002			Std B	L/N	16091005			Std C	L/N	16091006		
		Spec		Pass/Fail			Spec		Pass/Fail			Spec		Pass/Fail
		Na mmol/L	139.0-141.0	Pass			HCO3 mmol/L	39.2-40.8	Pass			pH @ 37C	7.352-7.358	Pass
		HCO3 mmol/L	19.6-20.4	Pass										
Std D	L/N	16091007			Reference	L/N	16090041							
		Spec		Pass/Fail			Spec		Pass/Fail					
		pH @ 37C	6.802-6.808	Pass			KCl mol/L	1.8-2.2	Pass					
		Na mmol/L	69.5-70.5	Pass										

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U. S. A. in accordance with ISO 9001:2008, Quality Management System Requirements and current Good Manufacturing Practices (cGMP) , and conforms to the indicated test specifications. The following analytes are traceable to NIST SRM Materials: Na, HCO3, pH, KCl.

Approval

QC Supervisor

Title

4/19/2016

Date