

CERTIFICATE OF ANALYSIS

Part Number 41874

Lot Number 16126024 Exp Date 11/03/2017 Std B L/N Std A L/N 16124003 16124004 Std C L/N 16124005 Pass/Fail Pass/Fail Pass/Fail Spec Spec Spec Na mmol/L 139.0-141.0 Pass HCO3 mmol/L 39.2-40.8 Pass pH @ 37C 7.352-7.358 Pass HCO3 mmol/L 19.6-20.4 Pass

Std D L/N 16124006 Reference L/N 16124002

Bio pHOx Cal Pack

 Spec
 Pass/Fail
 Spec
 Pass/Fail

 pH @ 37C 6.802-6.808
 Pass
 KCI mol/L
 1.8-2.2
 Pass

 Na mmol/L 69.5-70.5
 Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U. S. A. in accordance with ISO 9001:2008, Quality Management System Requirements and current Good Manufacturing Practices (cGMP), and conforms to the indicated test specifications. The following analytes are traceable to NIST SRM Materials: Na, HCO3, pH, KCI.

Approval

QC Supervisor

Title

5/19/2016 Date