




### CERTIFICATE OF ANALYSIS

STP pHox Plus C Cal Cartridge  
Lot Number 16155058

Exp Date 2017-11-23  
Date of Manufacture 2016-05-23  
Ref Number 41887

Cal A	L/N	16146035	Cal B	L/N	16153030	Cal C	L/N	16145032
	Spec	Pass/Fail		Spec	Pass/Fail		Spec	Pass/Fail
Na mmol/L	142-144	Pass	HCO3 mmol/L	44.10-45.90	Pass	pH @ 37C	7.327-7.333	Pass
HCO3 mmol/L	22.87-23.81	Pass				Na mmol/L	72.5-73.5	Pass
K mmol/L	4.03-4.09	Pass				K mmol/L	10.04-10.20	Pass
Cl mmol/L	97.2-98.8	Pass				Cl mmol/L	52.6-53.4	Pass
Ca mmol/L	1.01-1.03	Pass				Glu mg/dL	198-202	Pass
Glu mg/dL	39.6-40.4	Pass						
Cal D	L/N	16144063	Reference	L/N	16145042			
	Spec	Pass/Fail		Spec	Pass/Fail			
pH @ 37C	6.783-6.789	Pass	KCl	1.8-2.2	Pass			
Ca mmol/L	2.01-2.07	Pass						
Glu mg/dL	990-1010	Pass						

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U. S. A. in accordance with 21 CFR-Part 820, Quality System Regulation (QSR), Current Good Manufacturing Practice (cGMP) and EN ISO 13485:2012, Medical Devices, Quality Management Systems for in Vitro Diagnostic devices, and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

  
Approval

QC Supervisor  
Title

6/16/2016  
Date