




### CERTIFICATE OF ANALYSIS

STP pHox Plus C Cal Cartridge  
Lot Number 16155059

Ref Number 41889  
Exp Date 2017-11-23  
Date of Manufacture 2016-05-23

Cal A	L/N	16146035		Cal B	L/N	16153030		Cal C	L/N	16145032	
		Spec	Pass/Fail			Spec	Pass/Fail			Spec	Pass/Fail
Na mmol/L		142-144	Pass	HCO3 mmol/L		44.10-45.90	Pass	pH @ 37C		7.327-7.333	Pass
HCO3 mmol/L		22.87-23.81	Pass					Na mmol/L		72.5-73.5	Pass
K mmol/L		4.03-4.09	Pass					K mmol/L		10.04-10.20	Pass
Cl mmol/L		97.2-98.8	Pass					Cl mmol/L		52.6-53.4	Pass
Ca mmol/L		1.01-1.03	Pass					Glu mg/dL		198-202	Pass
Glu mg/dL		39.6-40.4	Pass								
Cal D	L/N	16144063		Reference	L/N	16145042					
		Spec	Pass/Fail			Spec	Pass/Fail				
pH @ 37C		6.783-6.789	Pass	KCl		1.8-2.2	Pass				
Ca mmol/L		2.01-2.07	Pass								
Glu mg/dL		990-1010	Pass								

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U. S. A. in accordance with 21 CFR-Part 820, Quality System Regulation (QSR), Current Good Manufacturing Practice (cGMP) and EN ISO 13485:2012, Medical Devices, Quality Management Systems for in Vitro Diagnostic devices, and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

  
Approval

QC Supervisor  
Title

6/16/2016  
Date