

CERTIFICATE OF ANALYSIS

Bioprofile Flex pH Sensor Lot Number 15294008 Part Number 42452

Exp Date 2017-10

Pass

Pass

Spec Pass/Fail

Slope 10.0-10.8 Std A mV 0+/-110

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U. S. A. in accordance with ISO 9001:2008, Quality Management System Requirements and current Good Manufacturing Practices (cGMP), and conforms to the indicated test specifications.

Joyce annaire

QC Supervisor

Title

11/5/2015

Date