



CERTIFICATE OF ANALYSIS

Bioprofile Flex pH Sensor
Lot Number 15322004

Part Number 42452
Exp Date 2017-11

	Spec	Pass/Fail
Slope	10.0-10.8	Pass
Std A mV	0+/-110	Pass

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U. S. A. in accordance with ISO 9001:2008, Quality Management System Requirements and current Good Manufacturing Practices (cGMP), and conforms to the indicated test specifications.

A handwritten signature in cursive script, appearing to read "Joyce Amara".

Approval

QC Supervisor

Title

12/10/2015

Date