



**CERTIFICATE OF ANALYSIS**

Bioprofile Flex PCO2 Electrode  
Lot Number 15279007

Part Number 42454  
Exp Date 2017-10

	Spec	Pass/Fail
Slope	9.5-11.3	Pass
Std A mV	0 +/- 150	Pass

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U. S. A. in accordance with ISO 9001:2008, Quality Management System Requirements and current Good Manufacturing Practices (cGMP), and conforms to the indicated test specifications.

Approval

QC Supervisor

Title

10/23/2015

Date