



CERTIFICATE OF ANALYSIS

Bioprofile Flex PO2 Electrode
Lot Number 15198027

Part Number 42456
Exp Date 2017-07

	Spec	Pass/Fail
Slope	-3.0 to -6.0	Pass

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U. S. A. in accordance with ISO 9001:2008, Quality Management System Requirements and current Good Manufacturing Practices (cGMP), and conforms to the indicated test specifications.

A handwritten signature in cursive script, appearing to read "Joyce Ansara".

Approval

QC Supervisor
Title

8/5/2015
Date

CONFIDENTIAL