



## CERTIFICATE OF ANALYSIS

Bioprofile Flex Ca Sensor

Part Number 42458

Lot Number 16105032

Exp Date 2017-07

	Spec*	Pass/Fail
Ca Slope	$\geq 9.3$	Pass
Std C mV	$0 \pm 60$	Pass
BioPro Ctl L3	2.25-2.85	Pass

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U. S. A. in accordance with ISO 9001:2008, Quality Management System Requirements and current Good Manufacturing Practices (cGMP), and conforms to the indicated test specifications.

\*Note: The process for assigning values to Nova Biomedical Quality Control and Linearity Solutions utilizes multiple sources of test data inputs to establish the ranges for each analyte in each lot, which may result in small lot-to-lot variation in insert range values.

A handwritten signature in dark ink, appearing to read "Joyce Ansara".

Approval

QC Supervisor

Title

5/12/2016

Date