



CERTIFICATE OF ANALYSIS

Bioprofile Flex K Sensor
Lot Number 16130020

Part Number 42460
Exp Date 2017-08

	Spec*	Pass/Fail
Slope	≥ 9.6	Pass
Std A mV	0 ± 60	Pass
pHOx Ctl L1	5.46-6.06	Pass
pHOx Ctl L3	1.84-2.24	Pass

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U. S. A. in accordance with ISO 9001:2008, Quality Management System Requirements and current Good Manufacturing Practices (cGMP), and conforms to the indicated test specifications.

*Note: The process for assigning values to Nova Biomedical Quality Control and Linearity Solutions utilizes multiple sources of test data inputs to establish the ranges for each analyte in each lot, which may result in small lot-to-lot variation in insert range values.

A handwritten signature in cursive script, appearing to read "Joyce Ansara".

Approval

QC Supervisor

Title

5/23/2016

Date