



CERTIFICATE OF ANALYSIS

Bioprofile Flex Na Sensor
Lot Number 16078024

Part Number 42462
Exp Date 2017-09

| | | |
|----------|----------|-----------|
| | Spec | Pass/Fail |
| Slope | 9.0-10.5 | Pass |
| Std A mV | 0+/-80 | Pass |

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U. S. A. in accordance with ISO 9001:2008, Quality Management System Requirements and current Good Manufacturing Practices (cGMP), and conforms to the indicated test specifications.

A handwritten signature in dark ink, appearing to read "Joyce Amara".

Approval

QC Supervisor

Title

4/8/2016

Date