



CERTIFICATE OF ANALYSIS

Bioprofile Flex Na Sensor
Lot Number 16124025

Part Number 42462
Exp Date 2017-11

	Spec	Pass/Fail
Slope	9.0-10.5	Pass
Std A mV	0+/-80	Pass

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U. S. A. in accordance with ISO 9001:2008, Quality Management System Requirements and current Good Manufacturing Practices (cGMP), and conforms to the indicated test specifications.

Approval

QC Supervisor

Title

5/17/2016

Date