



CERTIFICATE OF ANALYSIS

Bioprofile Flex Reference Electrode
Lot Number 16034022

Part Number 42466
Exp Date 2017-08

	Spec	Pass/Fail
Na Slope	9.0-11.0	Pass
ADT 4	>/= 220	Pass

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U. S. A. in accordance with ISO 9001:2008, Quality Management System Requirements and current Good Manufacturing Practices (cGMP), and conforms to the indicated test specifications.

A handwritten signature in dark ink, appearing to read "Joyce Ansara".

Approval

QC Supervisor

Title

2/25/2016

Date