



CERTIFICATE OF ANALYSIS

Nova Glu/Lac Sensor Soaking Solution
Lot Number 16263004

Ref Number 13409
Exp Date 2017-09-19
Date of Manufacture 2016-09-19

	Spec	Pass/Fail
NaOCl	0.43-0.73%	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U. S. A. in accordance with 21 CFR-Part 820, Quality System Regulation (QSR), Current Good Manufacturing Practice (cGMP) and EN ISO 13485:2012, Medical Devices, Quality Management Systems for in Vitro Diagnostic devices, and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

A handwritten signature in cursive script, appearing to read "Joyce Ansara".

Approval

QC Supervisor
Title

9/30/2016
Date