



CERTIFICATE OF ANALYSIS

Bioprofile Flex Glutamate Sensor
Lot Number 15244027

Part Number 42472
Exp Date 2017-09

	Spec*	Pass/Fail
Slope	11.0-45.0	Pass
pHOx Ctl L3	76-90	Pass
Set A L1	470-550	Pass

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U. S. A. in accordance with ISO 9001:2008, Quality Management System Requirements and current Good Manufacturing Practices (cGMP), and conforms to the indicated test specifications.

*Note: The process for assigning values to Nova Biomedical Quality Control and Linearity Solutions utilizes multiple sources of test data inputs to establish the ranges for each analyte in each lot, which may result in small lot-to-lot variation in insert range values.


Approval

QC Supervisor
Title

9/17/2015
Date

CONFIDENTIAL