



CERTIFICATE OF ANALYSIS

Nova 1/5/10/11/13 Microsample Std
Lot Number 16293012

Ref Number 13984
Exp Date 2017-10-19
Date of manufacture 2016-10-19

	Spec	Pass/Fail
Na mmol/L	27.6-28.4	Pass
K mmol/L	0.79-0.81	Pass
Cl mmol/L	22.6-23.3	Pass
Li mmol/L	0.098-0.102	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U. S. A. in accordance with 21 CFR-Part 820, Quality System Regulation (QSR), Current Good Manufacturing Practice (cGMP) and EN ISO 13485:2012, Medical Devices, Quality Management Systems for in Vitro Diagnostic devices, and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

A handwritten signature in cursive script, appearing to read "Joyce Anzara".

Approval

QC Supervisor

Title

10/27/2016

Date