



CERTIFICATE OF ANALYSIS

Nova 8 Calibrator Pack
Lot Number 608177

Ref Number 15196
Exp Date 2017-09

Cal A	L/N	16081077		Cal B	L/N	16081077		Cal C	L/N	16081077	
		Spec	Pass/Fail			Spec	Pass/Fail			Spec	Pass/Fail
Na mmol/L	139.0-141.0		Pass	Na mmol/L	74.5-75.5		Pass	Na mmol/L	139.0-141.0		Pass
K mmol/L	3.97-4.03		Pass	K mmol/L	9.8-10.2		Pass	K mmol/L	3.97-4.03		Pass
iCa mmol/L	0.99-1.01		Pass	iCa mmol/L	1.98-2.02		Pass	iCa mmol/L	0.99-1.01		Pass
iMg mmol/L	0.495-0.505		Pass	iMg mmol/L	0.495-0.505		Pass	iMg mmol/L	1.48-1.52		Pass
pH @ 37C	7.274-7.294		Pass	pH @ 37C	6.730-6.750		Pass	pH @ 37C	7.374-7.394		Pass

Reference	L/N	16081077	
		Spec	Pass/Fail
KCl mol/L	1.8-2.2		Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U. S. A. in accordance with 21 CFR-Part 820, Quality System Regulation (QSR), Current Good Manufacturing Practice (cGMP) and ISO EN 13485:2012, Medical Devices, Quality Management Systems for in Vitro Diagnostic devices, and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Approval

QC Supervisor
Title

4/13/2016
Date