



CERTIFICATE OF ANALYSIS

Bioprofile Control Level 5
Lot Number 612302

Catalog Number 24619
Exp Date 2017-11

	Spec	Pass/Fail
Gln mmol/L	4.60-5.40	Pass
Glu mmol/L	1.80-2.20	Pass
Gluc g/L	9.60-10.60	Pass
Lac g/L	4.60-5.40	Pass

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U. S. A. in accordance with ISO 9001:2008, Quality Management System Requirements and current Good Manufacturing Practices (cGMP), and conforms to the indicated test specifications. The following analytes are traceable to NIST SRM Materials: Lac, Gluc.

A handwritten signature in cursive script, appearing to read "Joyce Anzara".

Approval

QC Supervisor

Title

5/25/2016

Date