

## **CERTIFICATE OF ANALYSIS**

Nova 12 Calibrator Pack Lot Number 630152 Ref Number 15650 Exp Date 2017-07-27 Date of Manufacture 2016-10-27

K mmol/L CI mmol/L HCO3 mmol/L BUN mg/dl	16301052 Spec 139.0-141.0 3.97-4.03 112.1-113.9 19.2-20.8 9.7-10.3	Pass/Fail Pass Pass Pass Pass Pass Pass	Cal B L/N  Na mmol/L  K mmol/L  Cl mmol/L  HCO3 mmol/L  BUN mg/dl	9.8-10.2 57.4-58.6 9.8-10.4 49.4-50.6	Pass/Fail Pass Pass Pass Pass Pass	Cal C L/N Na mmol/L Cl mmol/L NH4		Pass/Fail Pass Pass Pass
Glu mg/dl Cal J L/N Na mmol/L Cl mmol/L NH4	79.2-83.2 16301052 Spec 98.0-102.0 90.7-94.4 4.4-4.6	Pass/Fail Pass Pass Pass Pass	Glu mg/dl Flush L/N Aqueous Salt Solution	16307054 Spec N/A	Pass Result Present	H L/N CO2 Releasing Agent	16301052 Spec N/A	Result Present

Reference L/N 16301052

Spec Pass/Fail KCl mol/L 1.8-2.2 Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U. S. A. in accordance with 21 CRF-Part 820, Quality System Regulation (QSR), Current Good Manufacturing Practice (cGMP) and EN ISO 13485:2012, Medical Devices, Quality Management Systems for in Vitro Diagnostic devices, and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Approval

QC Supervisor

Jayce ansara

Title

12/8/2016 Date