



### CERTIFICATE OF ANALYSIS

Bioprofile Control Level 7  
Lot Number 605360

Catalog Number 33837  
Exp Date 2017-08

	Spec	Pass/Fail
pH	6.92-6.98	Pass
Gluc g/L	9.40-10.60	Pass
Ace mmol/L	80.0-120.0	Pass
PO4 mmol/L	3.60-4.40	Pass
Gly g/L	4.50-5.50	Pass
NH4+ mmol/L	9.4-10.6	Pass
Na+ mmol/L	90-100	Pass
K+ mmol/L	9.2-10.8	Pass
Lac g/L	2.70-3.30	Pass

This product is certified to have been manufactured and tested in accordance with ISO 9001:2008, Quality Management System Requirements and current Good Manufacturing Practices (cGMP), and conforms to the indicated test specifications.

A handwritten signature in dark ink, appearing to read "Joyce Anwar".

Approval

QC Supervisor

Title

3/21/2016

Date