



### CERTIFICATE OF ANALYSIS

Bioprofile Control Level 6  
Lot Number 614181

Catalog Number 33836  
Exp Date 2017-11

	Spec	Pass/Fail
pH	7.530-7.590	Pass
Gluc g/L	1.80-2.20	Pass
Ace mmol/L	17.0-23.0	Pass
PO4 mmol/L	0.85-1.15	Pass
Gly g/L	2.20-2.80	Pass
NH4+ mmol/L	1.7-2.3	Pass
Na+ mmol/L	135-145	Pass
K+ mmol/L	3.6-4.4	Pass
Lac g/L	0.90-1.10	Pass

This product is certified to have been manufactured and tested in accordance with ISO 9001:2008, Quality Management System Requirements and current Good Manufacturing Practices (cGMP), and conforms to the indicated test specifications.

A handwritten signature in dark ink, appearing to read "Joyce Anwar".

Approval

QC Supervisor

Title

6/30/2016

Date