



### CERTIFICATE OF ANALYSIS

Bioprofile Control Level 4  
Lot Number 614643

Catalog Number 24617  
Exp Date 2017-11

	Spec	Pass/Fail
Gln mmol/L	0.85-1.15	Pass
Glu mmol/L	0.40-0.60	Pass
Gluc g/L	0.85-1.15	Pass
Lac g/L	0.85-1.15	Pass

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U. S. A. in accordance with ISO 9001:2008, Quality Management System Requirements and current Good Manufacturing Practices (cGMP), and conforms to the indicated test specifications. The following analytes are traceable to NIST SRM Materials: Lac, Gluc.

A handwritten signature in dark ink, appearing to read "Joyce Amara".

Approval

QC Supervisor

Title

6/16/2016

Date