



CERTIFICATE OF ANALYSIS

StatPro pHox/CCX SO2 Calibrator

Ref Number 22771

Level 1 L/N: 16005026			Level 2 L/N: 16012020		
Level 1 Exp Date: 2017-07			Level 2 Exp Date: 2017-07		
	Spec	Pass/Fail		Spec	Pass/Fail
SO2 %	45-65	Pass	SO2 %	92-105	Pass
Hct %	51-55	Pass	Hct %	51-55	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U. S. A. in accordance with 21 CFR-Part 820, Quality System Regulation (QSR), Current Good Manufacturing Practice (cGMP) and EN ISO 13485:2012, Medical Devices, Quality Management Systems for in Vitro Diagnostic devices, and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials. SO2 is traced to tonometry.



Approval

QC Supervisor
Title

3/21/2016
Date