



CERTIFICATE OF ANALYSIS

StatPro pHox/CCX SO2 Calibrator

Ref Number 22771

Lot Number 16159035

Exp Date 2017-11-14
Date of Manufacture 2016-05-14

Level 1 L/N:	16135001
Level 1 Exp Date:	2017-11
Spec	Pass/Fail
SO2 %	45-65 Pass
Hct %	51-55 Pass

Level 2 L/N:	16135002
Level 2 Exp Date:	2017-11
Spec	Pass/Fail
SO2 %	92-105 Pass
Hct %	51-55 Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U. S. A. in accordance with 21 CFR-Part 820, Quality System Regulation (QSR), Current Good Manufacturing Practice (cGMP) and EN ISO 13485:2012, Medical Devices, Quality Management Systems for in Vitro Diagnostic devices, and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials. SO2 is traced to tonometry.

Approval

QC Supervisor

Title

6/15/2016

Date