



CERTIFICATE OF ANALYSIS

STP pHox/Basic Control Levels 1, 2, 3 - Ampules

Ref Number 22763

Level 1 L/N:	15212006
Level 1 Exp Date:	2017-07
Spec	Pass/Fail
pH	7.185-7.225
pCO2 mmHg	57.0-67.0
pO2 mmHg	57.0-67.0
SO2 %	60.0-67.0
Hct %	32-36

Level 2 L/N:	15205026
Level 2 Exp Date:	2017-07
Spec	Pass/Fail
pH	7.430-7.470
pCO2 mmHg	37.0-43.0
pO2 mmHg	96.0-106.0

Level 3 L/N:	15205027
Level 3 Exp Date:	2017-07
Spec	Pass/Fail
pH	7.565-7.605
pCO2 mmHg	19.0-24.0
pO2 mmHg	131.0-145.0
SO2 %	88.0-96.0
Hct %	48-52

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U. S. A. in accordance with 21 CFR-Part 820, Quality System Regulation (QSR), Current Good Manufacturing Practice (cGMP) and EN ISO 13485:2012, Medical Devices, Quality Management Systems for in Vitro Diagnostic devices, and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials. SO2 is traced to tonometry.

Approval

QC Supervisor

Title

9/3/2015

Date

CONFIDENTIAL