



### CERTIFICATE OF ANALYSIS

Bioprofile Control Level 3  
Lot Number 610351

Catalog Number 22602  
Exp Date 2017-10

	Spec	Pass/Fail
pH	7.250-7.450	Pass
PO2 mmHg	128.0-173.0	Pass
PCO2 mmHg	16.0-24.0	Pass
Gluc g/L	14.20-15.80	Pass
Lac g/L	4.50-5.50	Pass
NH4+ mmol/L	9.25-10.75	Pass
Na+ mmol/L	176.0-186.0	Pass
K+ mmol/L	5.35-6.65	Pass
iCa++ mmol/L	2.40-3.00	Pass
Osmo mOsm/Kg	598-610	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U. S. A. in accordance with ISO 9001:2008, Quality Management System Requirements and current Good Manufacturing Practices (cGMP) , and conforms to the indicated test specifications. The following analytes are traceable to NIST SRM Materials: pH, PO2, PCO2, Lac, Gluc, NH4, Na, K, iCa.

A handwritten signature in dark ink, appearing to read "Joyce Amara", written in a cursive style.

Approval

QC Supervisor

Title

5/11/2016

Date