



## CERTIFICATE OF ANALYSIS

Bioprofile Control Level 2  
Lot Number 601145

Catalog Number 22601  
Exp Date 2017-07

	Spec	Pass/Fail
pH	6.830-7.030	Pass
PO2 mmHg	94.0-128.0	Pass
PCO2 mmHg	36.0-48.0	Pass
Gluc g/L	4.50-5.50	Pass
Lac g/L	2.70-3.30	Pass
NH4+ mmol/L	4.50-5.50	Pass
Na+ mmol/L	126.0-136.0	Pass
K+ mmol/L	3.75-4.25	Pass
iCa++ mmol/L	0.65-0.95	Pass
PO4 mmol/L	0.80-1.20	Pass
Osmo mOsm/Kg	392-400	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U. S. A. in accordance with ISO 9001:2008, Quality Management System Requirements and current Good Manufacturing Practices (cGMP), and conforms to the indicated test specifications. The following analytes are traceable to NIST SRM Materials: pH, PO2, PCO2, Lac, Gluc, NH4, Na, K, iCa.

Approval

QC Supervisor  
Title

3/10/2016  
Date