



### CERTIFICATE OF ANALYSIS

Bioprofile Control Level 1  
Lot Number 608870

Catalog Number 22600  
Exp Date 2017-09

	Spec	Pass/Fail
pH	6.580-6.780	Pass
PO2 mmHg	54.0-74.0	Pass
PCO2 mmHg	58.0-78.0	Pass
Gluc g/L	0.85-1.15	Pass
Lac g/L	0.85-1.15	Pass
NH4+ mmol/L	0.65-1.15	Pass
Na+ mmol/L	72.0-82.0	Pass
K+ mmol/L	1.85-2.15	Pass
iCa++ mmol/L	0.55-0.75	Pass
PO4 mmol/L	3.9-4.4	Pass
Osmo mOsm/Kg	213-219	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U. S. A. in accordance with ISO 9001:2008, Quality Management System Requirements and current Good Manufacturing Practices (cGMP) , and conforms to the indicated test specifications. The following analytes are traceable to NIST SRM Materials: pH, PO2, PCO2, Lac, Gluc, NH4, Na, K, iCa.

Approval

QC Supervisor

Title

4/26/2016

Date