



## CERTIFICATE OF ANALYSIS

Nova Control Levels 1 and 2

Ref Number 18655

Level 1 L/N	15310005		Level 2 L/N	15310006	
Level 1 Exp Date	2017-11		Level 2 Exp Date	2017-11	
	Spec	Pass/Fail		Spec	Pass/Fail
Na+ mmol/L	132-156	Pass	Na+ mmol/L	110-129	Pass
K+ mmol/L	3.2-4.0	Pass	K+ mmol/L	5.6-6.8	Pass
Cl- mmol/L	95-114	Pass	Cl- mmol/L	76-92	Pass
TCO2 mmol/L	19-30	Pass	TCO2 mmol/L	10-18	Pass
Glu mg/dL	69-91	Pass	Glu mg/dL	233-309	Pass
BUN mg/dL	15-25	Pass	BUN mg/dL	48-68	Pass
TCa mg/dL	5.04-7.24	Pass	TCa mg/dL	7.74-10.91	Pass
Ca++ mg/dL	3.96-5.72	Pass	Ca++ mg/dL	6.49-8.85	Pass
Mg++ mg/dL	0.98-1.74	Pass	Mg++ mg/dL	3.38-4.65	Pass
Li+ mmol/L	0.56-0.87	Pass	Li+ mmol/L	1.84-2.56	Pass
pH	7.30-7.40	Pass	pH	7.55-7.60	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U. S. A. in accordance with 21 CFR-Part 820, Quality System Regulation (QSR), Current Good Manufacturing Practice (cGMP) and EN ISO 13485:2012, Medical Devices, Quality Management Systems for in Vitro Diagnostic devices, and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Approval

QC Supervisor

Title

12/29/2015

Date