



CERTIFICATE OF ANALYSIS

PO2 Internal Filling Solution
Lot Number 16088014

Ref Number 06554
Exp Date 2017-09

	Spec	Pass/Fail
Na mM	60.0-80.0	Pass
K mM	27.0-35.0	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U. S. A. in accordance with 21 CFR-Part 820, Quality System Regulation (QSR), Current Good Manufacturing Practice (cGMP) and EN ISO 13485:2012, Medical Devices, Quality Management Systems for in Vitro Diagnostic devices, and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

A handwritten signature in dark ink, appearing to read "Joyce Anzani".

Approval

QC Supervisor
Title

5/10/2016
Date