




CERTIFICATE OF ANALYSIS

StatPro Control Levels 1/2

Ref Number 08161

Level 1 L/N: 14192014			Level 2 L/N: 14192015		
Level 1 Exp Date: 2017-07			Level 2 Exp Date: 2017-07		
	Spec	Pass/Fail		Spec	Pass/Fail
Na mmol/L	169.0-179.0	Pass	Na mmol/L	163.0-173.0	Pass
Hct %	31-35	Pass	Hct %	50-54	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U. S. A. in accordance with 21 CFR-Part 820, Quality System Regulation (QSR), Current Good Manufacturing Practice (cGMP) and EN ISO 13485:2012, Medical Devices, Quality Management Systems for in Vitro Diagnostic devices, and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.


Approval

QC Supervisor
Title

8/15/2014
Date

CONFIDENTIAL