



CERTIFICATE OF ANALYSIS

BioProfile pH_Ox Auto-Cartridge QC
Lot Number 16103050

Ref Number 42751
Exp Date 2017-09

	Level 1	16088096
	Spec	Pass/Fail
pH	6.600-6.800	Pass
PCO ₂ mmHg	57.0-67.0	Pass
PO ₂ mmHg	53.0-63.0	Pass

	Level 2	16090010
	Spec	Pass/Fail
pH	6.750-6.950	Pass
PCO ₂ mmHg	37.0-43.0	Pass
PO ₂ mmHg	93.0-103.0	Pass

	Level 3	16091008
	Spec	Pass/Fail
pH	7.200-7.400	Pass
PCO ₂ mmHg	17.5-22.5	Pass
PO ₂ mmHg	133.0-143.0	Pass

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U. S. A. in accordance with ISO 9001:2008, Quality Management System Requirements and current Good Manufacturing Practices (cGMP), and conforms to the indicated test specifications. The following analytes are traceable to NIST SRM Materials: pH, PCO₂, PO₂.

Approval

QC Supervisor

Title

4/26/2016

Date