



### CERTIFICATE OF ANALYSIS

BioPro Flex Deproteinizing Solution  
Lot Number 16266032

Ref Number 46850  
Exp Date 2017-09

	Spec	Pass/Fail
NaOCl %	0.7-0.9	Pass
pH	9.0-11.0	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U. S. A. in accordance with ISO 9001:2008, Quality Management System Requirements and current Good Manufacturing Practices (cGMP) , and conforms to the indicated test specifications. The following analytes are traceable to NIST SRM Materials: pH

A handwritten signature in cursive script, appearing to read "Joyce Anzani".

Approval

QC Supervisor

Title

9/30/2016

Date