




CERTIFICATE OF ANALYSIS

Bioprofile Flex Linearity Solution LOW
Lot Number 16126023

Catalog Number 45278
Exp Date 2017-11

	Spec	Pass/Fail
Gln mmol/L	0.35-0.65	Pass
Glu mmol/L	0.35-0.65	Pass
Gluc g/L	0.35-0.65	Pass
Lac g/L	0.35-0.65	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U. S. A. in accordance with ISO 9001:2008, Quality Management System Requirements and current Good Manufacturing Practices (cGMP), and conforms to the indicated test specifications. The following analytes are traceable to NIST SRM materials: Gluc, Lac


Approval

QC Supervisor
Title

6/15/2016
Date