



### CERTIFICATE OF ANALYSIS

Nova StatStrip Glucose and  $\beta$ -Ketone Control Solution Level 1  
Lot Number 0415121301

Ref Number 46947  
Exp Date 2017-11

	Spec	Mean Result
Glucose mg/dL	58-63	62
Ketone mmol/L	0.1-0.2	0.1

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U. S. A. in accordance with 21 CFR-Part 820, Quality System Regulation (QSR), Current Good Manufacturing Practice (cGMP) and EN ISO 13485:2012, Medical Devices, Quality Management Systems for in Vitro Diagnostic devices, and conforms to the indicated test specifications.

A handwritten signature in dark ink, appearing to read "Joyce Amara", written in a cursive style.

Approval

QC Supervisor

Title

5/28/2015

Date

CONFIDENTIAL