



CERTIFICATE OF ANALYSIS

Nova StatStrip Glucose Control Solution Level 1
Lot Number 0415065301

Ref Number 41741
Exp Date 2017-09

	Spec	Mean Result
Glucose mg/dL	48-78	57

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U. S. A. in accordance with 21 CFR-Part 820, Quality System Regulation (QSR), Current Good Manufacturing Practice (cGMP) and EN ISO 13485:2012, Medical Devices, Quality Management Systems for in Vitro Diagnostic devices, and conforms to the indicated test specifications.


Approval

QC Supervisor
Title

4/7/2015
Date

CONFIDENTIAL