




**CERTIFICATE OF ANALYSIS**

Nova StatStrip Glucose Control Solution Level 1  
Lot Number 0415147301

Ref Number 41741  
Exp Date 2017-11

	Spec	Mean Result
Glucose mg/dL	48-78	61

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U. S. A. in accordance with 21 CFR-Part 820, Quality System Regulation (QSR), Current Good Manufacturing Practice (cGMP) and EN ISO 13485:2012, Medical Devices, Quality Management Systems for in Vitro Diagnostic devices, and conforms to the indicated test specifications.

  
Approval

QC Supervisor  
Title

6/26/2015  
Date

CONFIDENTIAL