nova biomedical

CERTIFICATE OF ANALYSIS

StatStrip Glucose Control Level 3

Nova Reference 41743 Date of Manufacture

Lot Number 0421306303 Exp Date 2024-05-02

	Spec	Mean Result
Glucose mg/dL	275-325	297

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector
Title

11/20/2021
Date