



### CERTIFICATE OF ANALYSIS

Nova StatStrip Lactate Control Solution Level 2  
Lot Number 2715005302

Ref Number 47554  
Exp Date 2017-07

	Spec	Mean Result
Lactate mmol/L	5.7-7.1	6.5

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U. S. A. in accordance with 21 CFR-Part 820, Quality System Regulation (QSR), Current Good Manufacturing Practice (cGMP) and EN ISO 13485:2012, Medical Devices, Quality Management Systems for in Vitro Diagnostic devices, and conforms to the indicated test specifications.

A handwritten signature in dark ink, appearing to read "Joyce Amara".

Approval

QC Supervisor  
Title

2/17/2015  
Date

CONFIDENTIAL