



CERTIFICATE OF ANALYSIS

Bioprofile Flex Phosphate Reagent Pack				Part Number 39582	
Lot Number 17059024				Exp Date 2017-11	
Soln P L/N	17059024		Soln A L/N	17059024	
Phosphate mM	Spec 4.75-5.25	Result Pass	Ammonium Salt	Spec N/A	Result Present
Soln B L/N	17059024		Soln C L/N	17059024	
Reducing Agent	Spec N/A	Result Present	pH	Spec 6.900-7.100	Result Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U. S. A. in accordance with ISO 9001:2008, Quality Management System Requirements and current Good Manufacturing Practices (cGMP) , and conforms to the indicated test specifications. The following analyte is traceable to NIST SRM Materials: pH.

Approval

QC Supervisor

Title

3/16/2017

Date