



CERTIFICATE OF ANALYSIS

BioProfile CDV Reagent Pack			Ref Number 48796		
Lot Number 16279026			Exp Date 2017-09		
CDV Bottle Pack L/N: 16279028			Exp Date: 2017-09		
C Soln L/N	16256008		T Soln L/N	16267049	
	Spec	Result		Spec	Result
3% Detergent	N/A	Present	0.4% Trypan	N/A	Present
pH	>9.0	Pass	Particle Count	≤ 2.0	Pass
CDV Cartridge L/N: 16279027			Exp Date: 2017-09		
S Soln L/N	16258052		D Soln L/N	16257038	
	Spec	Pass/Fail		Spec	Result
pH	6.900-7.100	Pass	%NaOCl	0.43-0.73	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U. S. A. in accordance with ISO 9001:2008, Quality Management System Requirements and current Good Manufacturing Practices (cGMP) , and conforms to the indicated test specifications. The following analytes are traceable to NIST SRM Materials: pH.

Approval

QC Supervisor

Title

10/11/2016

Date