




CERTIFICATE OF ANALYSIS

Nova Na/pH Conditioning Solution
Lot Number 16253050

Ref Number 06856
Exp Date 2017-12-09
Date of Manufacture 2016-09-09

	Spec	Pass/Fail
pH	≤ 5.0	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U. S. A. in accordance with 21 CFR-Part 820, Quality System Regulation (QSR), Current Good Manufacturing Practice (cGMP) and EN ISO 13485:2012, Medical Devices, Quality Management Systems for in Vitro Diagnostic devices, and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.


Approval

QC Supervisor
Title

10/4/2016
Date