



CERTIFICATE OF ANALYSIS

Bioprofile pH Sensor
Lot Number 16131027

Part Number 23092
Exp Date 11/30/2017

	Spec	Pass/Fail
Slope	10.0-10.8	Pass
Std A mV	0 $\pm$ 110	Pass

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U. S. A. in accordance with ISO 9001:2008, Quality Management System Requirements and current Good Manufacturing Practices (cGMP), and conforms to the indicated test specifications.

A handwritten signature in cursive script, appearing to read "Joyce Ansara".

Approval

QC Supervisor
Title

5/25/2016
Date