



CERTIFICATE OF ANALYSIS

Bioprofile PCO2 Electrode
Lot Number 16144035

Part Number 23094
Exp Date 2017-07

	Spec	Pass/Fail
Slope	8.7-11.0	Pass
Std A mV	0 +/- 150	Pass

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U. S. A. in accordance with ISO 9001:2008, Quality Management System Requirements and current Good Manufacturing Practices (cGMP), and conforms to the indicated test specifications.

A handwritten signature in cursive script, appearing to read "Joyce Ansara".

Approval

QC Supervisor

Title

6/7/2016

Date